

# Z-Xtreme™ 5000T



## Stain-Resistant Labels for the Laboratory

### Facts

- Patient identification errors occur in about 1 percent of laboratory samples and approximately 1 in 18 sample identification errors lead to an adverse event.<sup>1</sup>
- Each year, there are 160,900 adverse events in U.S. hospitals because of sample identification errors.<sup>2</sup>
- These incidents have a significant financial impact, with redraws, retesting and additional treatment costing the healthcare industry \$200 to \$400 million annually.<sup>3</sup>

### Product Info

The durability of the labels utilized in the laboratory is a critical component of your patient safety system. If the label becomes unreadable, unscannable or falls off during testing, it can cause a sample identification error that could lead to an adverse event, re-testing, redraws and additional treatment.

Zebra's Z-Xtreme 5000T, coupled with our Image Lock™ ribbon, was designed to withstand the most common stains and chemicals found in laboratories—ensuring that the label does not discolor and the text and bar code remain readable and scannable during the label's useful life.



### Zebra's Z-Xtreme 5000T withstands exposure to:

- Xylene
- 4% Acetic Acid
- Ammonia
- 100% Isopropyl Alcohol
- Toluene
- Hexane
- Harris Hematoxylin
- Eosin
- And many others



Unlike standard synthetic labels (shown on left), Z-Xtreme 5000T (shown on right) is resistant to stains

In addition, this material is compatible with our G-Series™ desktop printers. These printers are small and fit easily into busy work environments.



There are several different technologies being used today in the laboratory for microscope slide labeling. As you can see in the chart to the right, the use of thermal labels is the most durable and the lowest cost.

	Technology	Etching	Thermal label	Handwritten
Equipment Cost		High	Low	Low
Label Cost		N/A	Low	N/A
Ease of Use		Quick and easy	Quick and easy	Time-consuming
Durability		High	High	High
Readability		Hard to read	Dark, crisp test	Depends on individual's handwriting
Scannability		High rate of inaccuracy	Quick and accurate	N/A

<sup>1</sup> Paul N. Valenstein, RL Sirota "Identification errors in pathology and laboratory medicine," Archives of Pathology and Laboratory Medicine, Vol. 129, No. 10, pp. 1228-1233.

<sup>2</sup> Paul N. Valenstein, MD, Stephen S. Raab, MD, Molly K. Walsh, PhD "Identification Errors Involving Clinical Laboratories: A College of American Pathologists Q-Probes Study of Patient and Specimen Identification Errors at 120 Institutions," Archives of Pathology and Laboratory Medicine: Vol. 130, No. 8, pp. 1106-1113.

<sup>3</sup> Ibid.



www.zebra.com

**Corporate Headquarters**  
+1 800 423 0442  
E-mail: inquiry4@zebra.com

**Asia-Pacific Headquarters**  
+65 6858 0722  
E-mail: apacchannelmarketing@zebra.com

**EMEA Headquarters**  
+44 (0)1628 556000  
E-mail: mseurope@zebra.com

**Latin America Headquarters**  
+1 847 955 2283  
E-mail: inquiry4@zebra.com

#### Other Locations

**USA:** California, Georgia, Rhode Island, Texas, Wisconsin **Europe:** France, Germany, Italy, Netherlands, Poland, Spain, Sweden **Asia Pacific:** Australia, China, India, Japan, South Korea **Latin America:** Argentina, Brazil, Florida (USA), Mexico **Africa/Middle East:** Russia, South Africa, United Arab Emirates

©2010 ZIH Corp. All product names and numbers are Zebra trademarks, and Zebra and the Zebra head graphic are registered trademarks of ZIH Corp. All rights reserved. All other trademarks are the property of their respective owners.

GSA#: GS-35F-0268N  
P1019020 Rev. 1 (10/10)





[www.L-TronDirect.com](http://www.L-TronDirect.com)

800-830-9523

[info@L-Tron.com](mailto:info@L-Tron.com)

596 Fishers Station Dr | Victor, NY | 14564 | Suite 1 A

[www.L-Tron.com](http://www.L-Tron.com)

